



The special laser cut design of the stent guarantees a precise handling and allows an accurate release without stent shortening. The transparency material of the delivery system allows a precise and reliable stent placement under endoscopic view.

The biliary duct stents are preload on the TTS (through-the-scope) delivery system and are placed through the working channel of the endoscope. The delivery system is introduced into the biliary duct through the papilla under visual and x-ray control.

- Exact positioning without stent shortening
- Defined high radial force at stent center, soft ends
- Laser cut stent design
- Memory effect of nitinol material – high radial force and flexibility
- Radiopaque tantalum markers
- Delivery system

Length 1800 mm, TTS
Ø 2,8 mm / 8.5 French
Accepts guide wire 0.035 inch

aixstent · biliary duct · laser cut design · non covered

| REF | Length | Length Cover | Ø Middle | Ø Ends |
|------------|--------|--------------|----------|--------|
| 381-10-040 | 40 | without | 10 | 10 |
| 381-10-060 | 60 | without | 10 | 10 |
| 381-10-080 | 80 | without | 10 | 10 |
| 381-10-100 | 100 | without | 10 | 10 |